

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 11 2011

DataMaster cdm S/N <u>130185</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>TODD BRUND</u>	ID# <u>7672</u> Date <u>7/7/11</u>
A Agency <u>NPS GLACIER Bay NP</u>	Phone # <u>907 697 2687</u>
Instrument Location <u>BARTLETT COVE RANGER STATION</u>	
B Alco S/N <u>2K</u> Target Value <u>0.80</u> High Pressure <u>800 psi</u>	
Alco Test Values <u>0.082</u> <u>0.082</u> 1st Alco 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

8MB
7/11/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 2011. NJB

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

DIAGNOSTIC CHECK

A circular notary seal for Alaska. The outer ring contains the text "STATE OF ALASKA" at the top and "ALASKA ROTH" at the bottom. The inner circle contains the text "NOTARY PUBLIC" and "Commission Expires 12/31/2018".

```

"44X": 147, "0123456789": 2000000000
HIJKL"HIJKLMNOPQRSTUVWXYZ" : abcdefghijklmnop
qrstuvwxyz{ }~

```